

## *EC Declaration of Conformity*

Boule Medical AB declare that the products below

*Type of product;*

*Cellcounter*

*Brand name or trade mark;*


*Medonic CellAnalyzer CA620/530 Vet*

are in conformity with the following directives and with harmonized standards:

<b>Directives</b>	<b>Standards</b>
<i>89/336/EEC Electro Magnetic Compatibility Directive (EMC) with amendment 92/31/EEC</i>	EN 61326:1997, with amendment EN 61326/A1 (1998)
<i>73/23/EEC Low Voltage Directive (LVD)</i>	EN 61010-1: 1993

Stockholm 2006-02-06

Date and place of issue



Eva Johansson  
*QA Manager*