


# Certificate of Analysis

<b>Product</b>	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
<b>Product No</b>	1070030 1070039
<b>Lot No</b>	19424570
<b>Manufacturing Date</b>	2019-10-02
<b>Expiration Date</b>	2020-10-30

<b>Parameter</b>	<b>Specification</b>
Volume 20 $\mu$ L	CV $\leq$ 1,0%
WBC background	$\leq$ 0,2 x 10 <sup>9</sup> /L
RBC background	$\leq$ 0,03 x 10 <sup>12</sup> /L
HGB background	$\leq$ 0,2 x g/dL
PLT background	$\leq$ 10 x 10 <sup>9</sup> /L

Actual value meets specification.

Date: 2020-02-05

  
Therese Hernström  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

