


Installation Instructions on M-series

A detailed description of reagent installation / reagent change can be found in the instrument User's Manual under section 2 Installation.



Intended use

The Boule Medonic M-series Diluent RFID and Lyse RFID reagents are intended for use with the Medonic M-series hematology analyzer systems.



Storage and stability

The Boule Medonic M-series reagents are stable at least up until the expiration date as shown on labels when stored at 4 to 35 °C. Open container stability is 3 months if storage recommendations are followed.

Instructions

1. For proper performance **DO NOT** mix Boule reagents with non-Boule products.
2. It is recommended that both the Diluent and the Lyse reagents are placed at the instrument level or below.
3. The interlocked reagent system displays indicator and warning messages to alert the operator when reagents are running low and need to be changed. When this occurs perform the following:
 - a. Select [MENU] tab and Press [REAGENT SETUP].
 - b. Press [ENTER NEW REAGENT] and then press [ENTER TAG].
 - c. Place the tag reader on the marked position on the reagent container you want to use. When the reagent tag has been read a screen will display that the tag has been accepted.
 - d. It is now possible to read another reagent [ENTER ANOTHER TAG] or exit to previous menu [EXIT].
 - e. Remove the cap and seal on the new reagent container.
 - f. Transfer the reagent level sensor from the used container to the new reagent container.
 - g. The analyzer is now ready to resume operation or analyze samples. No priming or fill cycle is necessary when putting on a new reagent container, if indicator and warning messages are followed.

You can always refer to your Medonic M-series User's manual for reagent priming or changing.

IMPORTANT! Remember to verify calibration prior to running patient samples e.g. using Boule control material.

Notes

- Prevent contamination of the reagents with dust, etc by use of the protective cap (delivered with the instrument).
- The cardboard box can be recycled after removal of the LPDE cubitainers. The LPDE container can be recycled after removal of any remaining reagent and a rinse with water.

Ordering information and service

For international customers contact your local distributor for orders and support.

REF	Description	Packaging
1504460	Medonic M-series Diluent, RFID	20 L
1504461	Medonic M-series Lyse, RFID	5 L
1504465	Medonic M-series DualPack, RFID	1 L Lyse / 5 L Diluent




Installation Instructions on M-series M32

A detailed description of reagent installation / reagent change can be found in the instrument User's Manual under section 2 – Installation and reagent Setup.



Intended use

The Medonic M-series M32 system is interlocked with specified Boule reagents, M-series Diluent RFID and M-series Lyse RFID for optimal performance.



Storage and stability

The Boule Medonic M-series reagents are stable at least up until the expiration date as shown on labels when stored at 4 to 35 °C. Open container stability is 3 months if storage recommendations are followed.

Instructions

1. For proper performance **DO NOT** mix Boule reagents with non-Boule products.
2. It is recommended that both the Diluent and the Lyse reagents are placed at the instrument level or below.
3. The interlocked reagent system displays indicator and warning messages to alert the operator when reagents are running low and need to be changed. When this occurs perform the following:
 - a. Select **Quick Functions Menu** and then select **Add Reagent**.
 - b. Scan in RFID tags on reagent box, and when all reagent RFID tags are entered a screen will display that reagent RFID tags have been accepted.
 - c. Select **Exit** to return to the **Quick Functions Menu**.
 - d. Remove the cap and seal on the new reagent container.
 - e. Transfer the reagent tube assembly from the used container to the new reagent container.
 - f. The analyzer is now ready to resume operation or analyze samples. No priming or fill cycle is necessary when putting on a new reagent container, if indicator and warning messages are followed.

You can always refer to your Medonic M-series M32 User's manual for reagent priming or changing.

IMPORTANT! Remember to verify calibration prior to running patient samples e.g. using Boule control material.

Notes

- Prevent contamination of the reagents with dust, etc by use of the protective cap (delivered with the instrument).
- The cardboard box can be recycled after removal of the LPDE cubitainers. The LPDE container can be recycled after removal of any remaining reagent and a rinse with water.

Ordering information and service

For international customers contact your local distributor for orders and support.

REF	Description	Packaging
1504460	Medonic M-series Diluent, RFID	20 L
1504461	Medonic M-series Lyse, RFID	5 L
1504465	Medonic M-series DualPack, RFID	1 L Lyse / 5 L Diluent

