

Certificate of Analysis

Product	MEDONIC M51-D DILUENT (20L)
Product No	1504510
Lot No	20181230
Manufacturing Date	2018-12-30
Expiration Date	2020-12-29

Test	Specification
pH	7,4 – 7,8
Osmolality (mOsm/kg)	300 – 320
Conductivity (mS/cm)	15,5 – 16,5
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2019-03-07



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

