

CERTIFICATE OF ANALYSIS

Product: Swelab™Alfalyse

Lot No: 2009-001


Product Part #: 501-207

Exp. Date: 2023-08-31

Manufacturing Date: 2020-09-23

Quality Testing Date: 2020-09-25

	<u>Result</u>	<u>Acceptable Range</u>
pH (units):	8.27	7.54 – 8.30
Osmolality (mOsm/kg):	320	295 – 330
Conductivity (µS/cm):	18.43	18.00 – 18.50
Functional Test	Pass	Pass



Jeanne Roscoe
Quality and Regulatory Manager

10/10/2020

Date