

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™ Alfalyse, 5L

**Lot No:** 2011-001

**Product Part #:** 501-207

**Exp. Date:** 2023-10-31

**Manufacturing Date:** 2020-11-03

**Quality Testing Date:** 2020-11-05

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.26	Pass
Osmolality (mOsm/kg):	295 – 330	319	Pass
Conductivity (mS/cm):	18.00– 18.50	18.40	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality Assurance Manager

6/1/2021  
\_\_\_\_\_  
Date

