

CERTIFICATE OF ANALYSIS

Product: Swelab™ AlfaLyse

Lot No: 1474

Product Part #: 501-210

Exp. Date: 2024-04-30

Manufacturing Date: 2021-05-26

Quality Testing Date: 2021-06-01

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.28	Pass
Osmolality (mOsm/kg):	295 – 330	318	Pass
Conductivity (mS/cm):	18.00 – 18.50	18.40	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Louise Esteva
Quality Assurance



Date

