


Certificate of Analysis


Product	MEDONIC M-SERIES DILUENT, RFID (20 L)
Product No	1504460
Lot No	2106-525
Manufacturing Date	2021-06-07
Expiration Date	2024-06-07

Test	Specification	Actual value
pH	6,90 – 7,10	
Osmolality (mOsm/kg)	290 – 310	
Conductivity (mS/cm)	17,50 – 18,00	
Particles (PLT/RBC channel)	≤ 5	

Date: 2021-06-15



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

