

# Certificate of Analysis

|                           |                           |
|---------------------------|---------------------------|
| <b>Product</b>            | SWELAB ALFADILUENT (20 L) |
| <b>Product No</b>         | 1504124                   |
| <b>Lot No</b>             | 2109-574                  |
| <b>Manufacturing Date</b> | 2021-09-13                |
| <b>Expiration Date</b>    | 2024-09-13                |


| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6,70 – 6,90   |
| Osmolality (mOsm/kg)        | 294 – 312     |
| Conductivity (mS/cm)        | 16,98 – 17,93 |
| Particles (PLT/RBC channel) | ≤ 5           |

Actual value meets specification.

Date: 2021-09-22



Lubna Elabbas  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

