

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™ AlfaDiluent

**Lot No:** 2107-002

**Product Part #:** 501-208


**Exp. Date:** 2024-06-30


**Manufacturing Date:** 2021-07-21

**Quality Testing Date:** 2021-07-26

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.80	Pass
Osmolality (mOsm/kg):	294 – 312	298	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.31	Pass
Background (x10 <sup>3</sup> /μL):	0.0 – 8.0	0.3	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality Assurance Manager

  
\_\_\_\_\_  
Date

