


Certificate of Analysis

Product	Exigo System Cleaner (1.9 L)
Product No	1504203
Lot No	2109-585
Manufacturing Date	2021-09-28
Expiration Date	2024-09-28


Test	Specification
pH	7,00 – 8,00
Conductivity (mS)	15,25 – 17,00
Particles (PLT/RBC channel)	≤ 30

Actual values meet specifications.

Date: 2021-10-05



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

