

# Certificate of Analysis

<b>Product</b>	BOULE EASYCLEANER (50 mL)
<b>Product No</b>	1504513
<b>Lot No</b>	2021051501
<b>Manufacturing date</b>	2021-05-15
<b>Expiration Date</b>	2023-05-14

---

<b>Test Hypochlorite</b>	<b>Specification</b>
Hypochlorite concentration	3.50 – 4.95 %
pH	12.60 – 13.20
Blank	PLT $\leq 10 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$


---

Actual values meet specifications.

Date: 2021-10-21



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

