

Certificate of Analysis

Product	SWELAB ALFALYSE, RFID (5 L)
Product No	1504463
Lot No	2110-597
Manufacturing Date	2021-10-18
Expiration Date	2024-10-18

Test	Specification
pH	7,80 – 8,30
Osmolality (mOsm/kg)	295 – 330
Conductivity (mS/cm)	17,75 – 18,25
CBC Performance *	Passes test
Particles (WBC channel)	≤ 0,2


* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2021-10-27



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

