

Certificate of Analysis

Product	SWELAB ALFADILUENT, RFID (20 L)
Product No	1504462
Lot No	2111-604
Manufacturing Date	2021-11-01
Expiration Date	2024-11-01


Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2021-11-09



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

