

# Certificate of Analysis

<b>Product</b>	MEDONIC M-SERIES DILUENT, RFID (20 L)
<b>Product No</b>	1504460
<b>Lot No</b>	2111-611
<b>Manufacturing Date</b>	2021-11-15
<b>Expiration Date</b>	2024-11-15


<b>Test</b>	<b>Specification</b>
pH	6,90 – 7,10
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,50 – 18,00
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2021-11-22



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

