

# Certificate of Analysis

|                           |                                       |
|---------------------------|---------------------------------------|
| <b>Product</b>            | MEDONIC M-SERIES DILUENT, RFID (20 L) |
| <b>Product No</b>         | 1504460                               |
| <b>Lot No</b>             | 2111-614                              |
| <b>Manufacturing Date</b> | 2021-11-18                            |
| <b>Expiration Date</b>    | 2024-11-18                            |

| <b>Test</b>                 | <b>Specification</b> |
|-----------------------------|----------------------|
| pH                          | 6,90 – 7,10          |
| Osmolality (mOsm/kg)        | 290 – 310            |
| Conductivity (mS/cm)        | 17,50 – 18,00        |
| Particles (PLT/RBC channel) | ≤ 5                  |

Actual value meets specification.

Date: 2021-11-25



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

