

Certificate of Analysis

| | |
|---------------------------|---|
| Product | Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs |
| Product No | 1070030 1070039 |
| Lot No | 214245M5 |
| Manufacturing Date | 2021-10-29 |
| Expiration Date | 2024-09-30 |

| Parameter | Specification |
|-------------------|-----------------------------------|
| Volume 20 μ L | CV \leq 1,0% |
| WBC background | \leq 0,2 x 10 ⁹ /L |
| RBC background | \leq 0,03 x 10 ¹² /L |
| HGB background | \leq 0,2 x g/dL |
| PLT background | \leq 10 x 10 ⁹ /L |

Actual value meets specification.



Date: 2021-12-03


Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.