

Certificate of Analysis


| | |
|---------------------------|-----------------------------------------------------------------------------------------|
| Product | Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs |
| Product No | 1070030 1070039 |
| Lot No | 214245M6 |
| Manufacturing Date | 2021-11-08 |
| Expiration Date | 2024-09-30 |

| Parameter | Specification |
|-------------------|------------------------------|
| Volume 20 μ L | $CV \leq 1,0\%$ |
| WBC background | $\leq 0,2 \times 10^9/L$ |
| RBC background | $\leq 0,03 \times 10^{12}/L$ |
| HGB background | $\leq 0,2 \times g/dL$ |
| PLT background | $\leq 10 \times 10^9/L$ |

Actual value meets specification.



Date: 2021-11-25


Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

