

# Certificate of Analysis

<b>Product</b>	EXIGO DILUENT
<b>Product No</b>	1504328 (10 L)
<b>Lot No</b>	2111-623
<b>Manufacturing Date</b>	2021-11-30
<b>Expiration Date</b>	2024-11-30


<b>Test</b>	<b>Specification</b>
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5
Total bacterial growth (CFU/100ml)	≤ 5

Actual values meet specifications.

Date: 2021-12-07



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

