

Certificate of Analysis

Product	EXIGO LYSE (1.9 L)
Product No	1504202
Lot No	2112-626
Manufacturing Date	2021-12-04
Expiration Date	2024-12-04

Test	Specification
pH	6.95 – 7.25
Osmolality (mOsm/kg)	195 - 225
Conductivity (mS/cm)	12.10 – 12.80
CBC Performance *	Passes test
Particles (PLT/RBC channel)	≤ 5

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2021-12-14



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

