

# Certificate of Analysis

|                           |                          |
|---------------------------|--------------------------|
| <b>Product</b>            | EXIGO LYSE, RFID (1.9 L) |
| <b>Product No</b>         | 1504500                  |
| <b>Lot No</b>             | 2112-626                 |
| <b>Manufacturing Date</b> | 2021-12-04               |
| <b>Expiration Date</b>    | 2024-12-04               |

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6.95 – 7.25   |
| Osmolality (mOsm/kg)        | 195 - 225     |
| Conductivity (mS/cm)        | 12.10 – 12.80 |
| CBC Performance *           | Passes test   |
| Particles (PLT/RBC channel) | ≤ 5           |


\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date:2021-12-14



Lubna Elabbas  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

