

Certificate of Analysis


Product	SWELAB ALFADILUENT (20 L)
Product No	1504124
Lot No	2112-635
Manufacturing Date	2021-12-20
Expiration Date	2024-12-20

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.



Date: 2021-12-27



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

