

# Certificate of Analysis


<b>Product</b>	SWELAB ALFADILUENT, RFID (20 L)
<b>Product No</b>	1504462
<b>Lot No</b>	2112-636
<b>Manufacturing Date</b>	2021-12-21
<b>Expiration Date</b>	2024-12-21

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.



Date: 2021-12-27

  
Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

