

Certificate of Analysis

Product	BOULE EASYCLEANER (50 mL)
Product No	1504513
Lot No	2021110701
Manufacturing date	2021-11-07
Expiration Date	2023-11-06

Test Hypochlorite	Specification
Hypochlorite concentration	3.50 – 4.95 %
pH	12.60 – 13.20
Blank	PLT $\leq 10 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$

Actual values meet specifications.

Date: 2021-12-07



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

