

Certificate of Analysis

Product	EXIGO DILUENT
Product No	1504328 (10 L)
Lot No	2201-644
Manufacturing Date	2022-01-11
Expiration Date	2025-01-11

Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5
Total bacterial growth (CFU/100ml)	≤ 5

Actual values meet specifications.

Date: 2022-01-17



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

