

Certificate of Analysis

Product	Exigo System Cleaner, RFID (1.9 L)
Product No	1504503
Lot No	2201-648
Manufacturing Date	2022-01-13
Expiration Date	2025-01-13


Test	Specification
pH	7,00 – 8,00
Conductivity (mS)	15,25 – 17,00
Particles (PLT/RBC channel)	≤ 30

Actual values meet specifications.

Date: 2022-01-20



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

