a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Stopper

Lot No:

2201-202

Product Part #: 501-350

Exp. Date:

2024-12-31

Manufacturing Date:

2022-01-31

Quality Testing Date: 2022-02-07

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 – 11.20	11.17	Pass
Osmolality (mOsm/kg):	830 – 860	835	Pass
Conductivity (mS/cm):	42.7 – 43.3	43.1	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Quality and Regulatory Manager