

# Certificate of Analysis

<b>Product</b>	MEDONIC M-SERIES DUALPACK (1 L Lyse / 5 L Diluent)
<b>Product No</b>	1504128
<b>Lot No</b>	2201-646
<b>Manufacturing Date</b>	2022-01-10
<b>Expiration Date</b>	2025-01-10

Test	Specification Diluent	Specification Lyse
pH	6,90 – 7,10	6,80 – 7,20
Osmolality (mOsm/kg)	290 – 310	220 – 240
Conductivity (mS/cm)	17,50 – 18,00	14,00 – 15,00
CBC Performance *	Passes test	Passes test
Particles (PLT/WBC channels)	≤ 5	≤ 0,2


\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2022-02-02



Lubna Elabbas  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

