


Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	214545V7
Manufacturing Date	2022-01-04
Expiration Date	2024-11-30

Parameter	Specification
Volume 20 µL	CV ≤ 1,0%
WBC background	≤ 0,2 x 10 ⁹ /L
RBC background	≤ 0,03 x 10 ¹² /L
HGB background	≤ 0,2 x g/dL
PLT background	≤ 10 x 10 ⁹ /L

Actual value meets specification.



Date: 2022-02-02

Susanne el Makdisi

Susanne el Makdisi
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

