

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Diluent

Lot No: 2201-004

Product Part #: 501-211

Exp. Date: 2024-12-31

Manufacturing Date: 2022-01-21


Quality Testing Date: 2022-01-27

Parameter	Specification	Result	Pass/Fail
pH (units):	6.90 – 7.20	7.13	Pass
Osmolality (mOsm/kg):	290 – 310	291	Pass
Conductivity (mS/cm):	17.50 – 18.00	17.51	Pass
Background Test (x10 ³ /μL):	0.0 – 5.0	0.3	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Quality and Regulatory Manager



Date

