

Certificate of Analysis

| | |
|---------------------------|---------------------------------|
| Product | MEDONIC M-SERIES DILUENT (20 L) |
| Product No | 1504122 |
| Lot No | 2202-666 |
| Manufacturing Date | 2022-02-15 |
| Expiration Date | 2025-02-15 |


| Test | Specification |
|-----------------------------|----------------------|
| pH | 6,90 – 7,10 |
| Osmolality (mOsm/kg) | 290 – 310 |
| Conductivity (mS/cm) | 17,50 – 18,00 |
| Particles (PLT/RBC channel) | ≤ 5 |

Actual value meets specification.

Date: 2022-02-22



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

