

# Certificate of Analysis

<b>Product</b>	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
<b>Product No</b>	1070030 1070039
<b>Lot No</b>	21554533
<b>Manufacturing Date</b>	2022-01-31
<b>Expiration Date</b>	2024-12-31


<b>Parameter</b>	<b>Specification</b>
Volume 20 $\mu$ L	$CV \leq 1,0\%$
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times g/dL$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2022-02-22



Caroline Sparf  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

