


Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	21554534
Manufacturing Date	2022-01-31
Expiration Date	2024-12-31


Parameter	Specification
Volume 20 µL	CV ≤ 1,0%
WBC background	≤ 0,2 x 10 ⁹ /L
RBC background	≤ 0,03 x 10 ¹² /L
HGB background	≤ 0,2 x g/dL
PLT background	≤ 10 x 10 ⁹ /L

Actual value meets specification.

Date: 2022-02-22



Caroline Sparf
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

