

## CERTIFICATE OF ANALYSIS

**Product:** Medonic™ M-Series Lyse

**Lot No:** 2201-002

**Product Part #:** 501-212


**Exp. Date:** 2024-12-31


**Manufacturing Date:** 2022-01-25

**Quality Testing Date:** 2022-01-28

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.11	Pass
Osmolality (mOsm/kg):	220 – 240	232	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.16	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Regulatory and Compliance Manager

  
\_\_\_\_\_  
Date

