

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™ Alfalyse, 5L

**Lot No:** 2203-002

**Product Part #:** 501-207

**Exp. Date:** 2025-02-28

**Manufacturing Date:** 2022-03-08

**Quality Testing Date:** 2022-03-11

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.28	Pass
Osmolality (mOsm/kg):	295 – 330	316	Pass
Conductivity (mS/cm):	18.00– 18.50	18.32	Pass
Functional Test	Pass		Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality Assurance Manager

3/16/2022  
Date

