

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™AlfaDiluent

**Lot No:** 2203-003

**Product Part #:** 501-208

**Exp. Date:** 2025-02-28

**Manufacturing Date:** 2022-25-03

**Quality Testing Date:** 2022-29-03

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.79	Pass
Osmolality (mOsm/kg):	294 – 312	297	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.25	Pass
Background (x10 <sup>3</sup> /μL):	0.0 – 8.0	0.5	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe  
Quality Assurance Manager

3/30/2022  
Date

