

Certificate of Analysis

| | |
|---------------------------|---------------------------------|
| Product | SWELAB ALFADILUENT, RFID (20 L) |
| Product No | 1504462 |
| Lot No | 2204-697 |
| Manufacturing Date | 2022-04-03 |
| Expiration Date | 2025-04-03 |


| Test | Specification |
|-----------------------------|---------------|
| pH | 6,70 – 6,90 |
| Osmolality (mOsm/kg) | 294 – 312 |
| Conductivity (mS/cm) | 16,98 – 17,93 |
| Particles (PLT/RBC channel) | ≤ 5 |

Actual value meets specification.

Date: 2022-04-13



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

