


Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	220945E8
Manufacturing Date	2022-03-23
Expiration Date	2025-02-28

Parameter	Specification
Volume 20 µL	$CV \leq 1,0\%$
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \text{ x g/dL}$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.



Date: 2022-04-19



Lubna Elabbas
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

