

# Certificate of Analysis

<b>Product</b>	SWELAB ALFADILUENT (20 L)
<b>Product No</b>	1504124
<b>Lot No</b>	2204-704
<b>Manufacturing Date</b>	2022-04-18
<b>Expiration Date</b>	2025-04-18


Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2022-04-26



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

