

# Certificate of Analysis

**Product** BOULE EASYCLEANER (50 mL)  
**Product No** 1504513  
**Lot No.** 2021092401  
**Manufacturing date** 2021-09-24  
**Expiration Date** 2023-09-23

---

Test Hypochlorite	Specification
Hypochlorite concentration	3.50 – <5 %
pH	12.60 – 13.20
Blank	PLT $\leq 10 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$


---

Actual values meet specifications.

Date: 2022-04-21



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

