

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Lyse

Lot No: 2204-003

Product Part #: 501-212

Exp. Date: 2025-03-31

Manufacturing Date: 2022-04-13


Quality Testing Date: 2022-04-20

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.13	Pass
Osmolality (mOsm/kg):	220 – 240	231	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.15	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Regulatory and Compliance Manager



Date

