

Certificate of Analysis

Product	MEDONIC M-SERIES DILUENT, RFID (20 L)
Product No	1504460
Lot No	2205-715
Manufacturing Date	2022-05-17
Expiration Date	2025-05-17

Test	Specification
pH	6,90 – 7,10
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,50 – 18,00
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.



Date: 2022-05-24



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

