

Certificate of Analysis

| | |
|---------------------------|---------------------------------------|
| Product | MEDONIC M-SERIES DILUENT, RFID (20 L) |
| Product No | 1504460 |
| Lot No | 2205-718 |
| Manufacturing Date | 2022-05-24 |
| Expiration Date | 2025-05-24 |


| Test | Specification |
|-----------------------------|---------------|
| pH | 6,90 – 7,10 |
| Osmolality (mOsm/kg) | 290 – 310 |
| Conductivity (mS/cm) | 17,50 – 18,00 |
| Particles (PLT/RBC channel) | ≤ 5 |

Actual value meets specification.

Date: 2022-05-31



Caroline Sparf
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

