


Certificate of Analysis


Product	SWELAB ALFADILUENT, RFID (20 L)
Product No	1504462
Lot No	2206-722
Manufacturing Date	2022-06-02
Expiration Date	2025-06-02

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.



Date: 2022-06-08



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

