


Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	222145Q4
Manufacturing Date	2022-05-31
Expiration Date	2025-04-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background	\leq 0,2 x 10 ⁹ /L
RBC background	\leq 0,03 x 10 ¹² /L
HGB background	\leq 0,2 x g/dL
PLT background	\leq 10 x 10 ⁹ /L

Actual value meets specification.



Date: 2022-06-07



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

