


Certificate of Analysis

Product SWELAB ALFALYSE (5 L)
Product No 1504125
Lot No 2206-741
Manufacturing Date 2022-06-28
Expiration Date 2025-06-28

| Test | Specification |
|-------------------------|---------------|
| pH | 7,80 – 8,30 |
| Osmolality (mOsm/kg) | 295 – 330 |
| Conductivity (mS/cm) | 17,75 – 18,25 |
| CBC Performance * | Passes test |
| Particles (WBC channel) | ≤ 0,2 |

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.



Date: 2022-07-07



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

