

Certificate of Analysis

Product MEDONIC M-SERIES LYSE (5 L)

Product No 1504123

Lot No 2207-745

Manufacturing Date 2022-07-03

Expiration Date 2025-07-03

Test Specification

pH 6,80-7,20

Osmolality (mOsm/kg) 220 - 240

Conductivity (mS/cm) 14,00-15,00

CBC Performance test* Passes test

Particles (WBC channel) ≤ 0.2

Actual value meets specification.

Date: 2022-07-11

Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated. All information in this document is property of Boule Medical AB.



CE

^{*} Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.