Certificate of Analysis

Product	MEDONIC M51-D DILUENT (20L)
Product No	1504510
Lot No	2022041801
Manufacturing Date	2022-04-18
Expiration Date	2024-04-17

Test	**Specification**
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10^9/L)	≤ 5

Actual values meet specification.

BOULE MEDICAL AB Boule STOCKHOLM SWEDEN

Date: 2022-07-28

Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

Boule Medical AB
Domnarvsgatan 4
SE-163 53 Spånga, Sweden
E-mail: info@boule.com
www.boule.com