

Certificate of Analysis


| | |
|---------------------------|---|
| Product | Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs |
| Product No | 1070030 1070039 |
| Lot No | 222545V1 |
| Manufacturing Date | 2022-07-01 |
| Expiration Date | 2025-05-28 |

| Parameter | Specification |
|------------------|------------------------------|
| Volume 20 µL | CV ≤ 1,0% |
| WBC background | ≤ 0,2 x 10 ⁹ /L |
| RBC background | ≤ 0,03 x 10 ¹² /L |
| HGB background | ≤ 0,2 x g/dL |
| PLT background | ≤ 10 x 10 ⁹ /L |

Actual value meets specification.

Date: 2022-08-02


Sofia Karlström
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

