

Certificate of Analysis


Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	22274520
Manufacturing Date	2022-07-20
Expiration Date	2025-06-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times \text{g/dL}$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2022-08-02


Sofia Karlström
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

